

LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Rajathoppu  
 Lab No: 12102  
 Month: August  
 Year: 2023  
 Constituent Group: Chemistry III

- [Details About Robust Analysis](#)
- [Detail About Monthly Summary](#)
- [Detail about SDI](#)
- [All Analyser Result](#)
- [Print](#)

Date of Result Entered : 21/08/2023  
 Date of Report Published : 01/09/2023

Click on the analyte to view Graphical Data

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No. of Participants	DV	CV	SD	Year Value	SDI	U
1	GLUCOSE III	GODD-POD III	Any Analyser (Automation / Semi Automation)	64	174.34	5.46	9.51	159.1 mg/dL	-1.60	2.38
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation)	62	1.76	7.18	0.13	1.7 mg/dL	-0.48	0.03
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	149	113.96	4.36	4.97	117.9 mg/dL	0.79	0.81
<b>SDI Range</b> Within -1.00 to +1.00: Excellent. Within ±1.01 to ±2.00: Good. Within ±2.01 to ±3.99: Accept with caution. Warning Signal. Beyond ±3.0: Unacceptable performance. Action Signal.										

Links