



CHRISTIAN MEDICAL COLLEGE
DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - AUGUST 2023



Lab Name: Govt. Block PHC, Pudunagar
 Constituent Group: Chemistry III
 PT Item: Lyophilized human serum based
 Lab No: 7181
 Date of Result Entered: 20/08/2023
 Date of Report Published: 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No. of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	180 mg/dL	0.48	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.6 mg/dL	-1.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	107 mg/dL	-1.47	0.22
		SDI Range	Interpretation							
		Within -1.00 to +1.00	Excellent.							
		Within ±1.01 to ±2.00	Good.							
		Within ±2.01 to ±2.99	Accept with caution. Warning Signal.							
		Beyond ±3.0	Unacceptable performance. Action Signal.							

LAB ADDRESS:
 Govt. Block PHC, Pudunagar
 The Block Medical Officer, Pudunagar, Pudukottal District
 Pudunagar
 TAMILNADU613301

Coordinator Contact Details:
 Email: clinqt@cmcvellore.ac.in
 Contact Number: 0416-2283102

Paula Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore
 ARATHAKOTTAI (DO) -611301

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****