

LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Rajathoppu  
 Lab No: 12102  
 Month: August  
 Year: 2023  
 Constituent Group: Chemistry III

Click on the analyte to view Graphical Data

[All Analyser Result](#)    [Print](#)  
 Date of Result Entered : 21/08/2023  
 Date of Report Published : 01/09/2023

[Details About Robust Analysis](#)  
[Detail About Monthly Summary](#)  
[Detail about SDI](#)

Analyte	Method / Principle Name	Analyzer Name	No. of Participants	DV	CV	SD	Year Value	SDI	U
1 GLUCOSE III	GODD-POD III	Any Analyser (Automation / Semi Automation)	64	174.34	5.46	9.51	159.1 mg/dL	-1.60	2.38
2 CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation)	62	1.76	7.18	0.13	1.7 mg/dL	-0.48	0.03
3 CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	149	113.96	4.36	4.97	117.9 mg/dL	0.79	0.81
<b>SDI Range Interpretation</b>									
Within -1.00 to +1.00			Excellent.						
Within ±1.01 to ±2.00			Good.						
Within ±2.01 to ±3.00			Accept with caution. Warning Signal.						
Beyond ±3.0			Unacceptable performance. Action Signal.						

Links



Govt. PHC, Rajathoppu

Chemistry III

Lyophilized human serum based

# CHRISTIAN MEDICAL COLLEGE

## DEPARTMENT OF CLINICAL BIOCHEMISTRY CMC EXTERNAL QUALITY ASSURANCE SCHEME MONTHLY SUMMARY REPORT - JULY 2023



PC-1024

Lab No

12102

Date of Result Entered :

20/07/2023

Date of Report Published :

07/08/2023

Analyte	Method / Principle Name	Analyzer Name	No of Participants	Mean	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	75	109.71	5.43	106.4 mg/dL	-0.56	1.37
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	66	5.89	4.60	5.78 mg/dL	-0.41	0.07
3	CHOLESTROL III	CHOD-PAP III	148	112.78	4.33	99.1 mg/dL	-2.80	0.80
SDI Range				Interpretation				
Within -1.00 to +1.00				Excellent.				
Within ±1.01 to ±2.00				Good.				
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.				
Beyond ±3.0				Unacceptable performance. Action Signal.				

LAB ADDRESS :  
 Govt. PHC, Rajathoppu  
 THE MEDICAL OFFICER, RAJATHOPPU PHC, DHARMAPURI, DHARMAPURI DIST.  
 Rajathoppu  
 TAMILNADU636704

Coordinator Contact Details:  
 Email: cilmqcc@cmcvellore.ac.in  
 Contact Number: 0416-2283102

*Pamela Christudoss*  
 Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS report are confidential