



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - AUGUST 2023



Lab Name	Govt. Block PHC, Kallandiri	Lab No	3948
Constituent Group	Chemistry III	Date of Result Entered :	19/08/2023
PT item	Lyophilized human serum based	Date of Report Published :	01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Transasia / Erba	23	173.17	4.83	8.36	162.8 mg/dL	-1.24	3.49
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Transasia / Erba	21	1.74	6.34	0.11	1.8 mg/dL	0.55	0.05
3	CHOLESTROL III	CHOD-PAP III	Transasia / Erba	17	111.42	3.18	3.54	113.2 mg/dL	0.50	1.72

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. Block PHC, Kallandiri
 "The Block Medical Officer KALLANDIRI,
 Kallandiri
 TAMILNADU625301

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****