



CALIBRATION CERTIFICATE

Page no: 1 of 1

Customer Name & Address: M/S GOVERNMENT PRIMARY HEALTH CENTRE , THIRUVATTAR - 629 183.	Certificate No:	VCPL/CC/23-24/16-003
	Date of Calibration:	28-09-2023
	Date of issue:	29-09-2023
	Date of Receipt:	28-09-2023
	Next cal Date:	27-09-2024

Details of Device Under Calibration			
Description:	MICROPIPETTE	Make:	DRAGONLAB
Range:	(100 to 1000) μ l	Model/Type:	-
Resolution:	1 μ l	Condition of DUC	GOOD
Serial Number	YB17AAGOO41444	Location:	LABORATORY
ID number	-	Cal. At:	ONSITE
Environmental Conditions		Calibration Method Used	
Temperature	23.6 $^{\circ}$ C	Calibration Method Used	ISO 9000:2002 PART-8
Humidity	49 % Rh	Cal procedure no:	-

Master Details

SL no:	Description	ID No / SL No	Certificate No	Make/Model	Valid Till
1	ELECTRONIC BALANCE	16031172	E-150323-1	WINTech	15-03-2024

RESULT OF CALIBRATION

S.NO	Duc Readings (μ l)	Standard Reading Mean Value (mg)	ERROR (mg)
1	100	99.93	0.07
2	500	499.88	0.12
3	1000	999.82	0.18


Remarks:

- 1.DUC is Defined as Device Under Calibration.
- 2.Calibration Frequency decided by the customer.
- 3.The above instrument is working in good condition.


Calibrated By:


GAKASH

(CALIBRATION ENGINEER)



Authorized by:


S. LAKSHMANAN MURUGAN,
Technical Manager
Vanavil Calibrations Pvt Ltd.

Vanavil Calibrations Pvt. Ltd

(A Unit of Rainbow Bio Medical Engineers)

S2, Rishi Maharaj Apartment, Indra Nagar, 250 Reddiyur Main Road, Narasothipatty, Salem - 636 004.

o. 96,97,98, Anadhamaya Apartments, 2nd Floor, KVR Flats, 2nd Cross Street, Thangam Nagar, Pallikaranai, Chennai - 600 100.

Cell: 99949 13872 / 99949 23872 Ph: 0427-7965376

vanavilcalibrations@gmail.com www.rainbowbiomedicalengineers.com | www.vanavilcalibrations.com