

	ROOTS INDUSTRIES INDIA LIMITED ROOTS METROLOGY & TESTING LABORATORY R K G Industrial Estate, Ganapathy, Coimbatore-641 006, India Phone: 9585840192/9003432356	 NABL/ILAC/024 NABL ACCREDITATION CERTIFICATE NO. CC 2752
---	---	--

CALIBRATION CERTIFICATE

ULR Number: CC275219000083790F

Name and address of the customer The Medical Officer Government Primary Health Centre Kothamangalam -pudukottai(dist)	Certificate no : Rmtl/01/319107849 Date of calibration :11/04/2023 Suggested calibration due :10/04/2024 Customer po .reference.& date:
--	--

1.DESCRPTION OF DUC

Name Of The Duc: CENTRIFUGE Make :labotech Model : bdi 152 Sl.No : Range/Size : 1-4000RPM/230v	Receiving Condition Of Duc :Good Ref. Document :As Per Manufactures Specification Environmental Condition :25.3 Deg & 54.3% Rh Uncertainty Of Measurement :As Per Report Location :lab
---	---

THE MASTERS ARE TRACEABLE TO NATIONAL/INTERNATIONAL STANDARDS AS PER ISO/IEC 17025.

2. DETAILS OF MASTER USED

Standard Instrument Used	Make/ Model	Id No	Traceable To/Certificate No	Validity
Decade Megohm	Vaisashika	Rml-230	Vaiseshika,Ambala Cantt/Ved/Nabl/18-19(Et)/92	1-6-2024
Digital Tachometer	Mextec	Rml-251	Gsg/19/206/Tm/01	1-6-2024
Multi Product Calibrator	Fluke	Rml-262	Wavecal,Blr/Cs-18/01-01	1-6-2024
Multi Product Calibrator (Ac/Dc Current Coil)	Transmille	Rml-252-1	Wavecal,Blr/Cs-18/69-01	1-6-2024
Leakage Current Tester	Htc/CI-2054	Rml-254-1	Rmtl/01/319107848	1-6-2024

CALIBRATION FIELD: Electro Technical (Voltage Calibration -Ac)

Sl.no	Range	Master readings v	Calibrated values v	Error observed v	Uncertainty v
1	0 v to 400v	230	231	1	0.055

CURRENT CALIBRATIONAc Current @50Hz

Sl.no	Range	Master readings mA	Calibrated values mA	Error observed mA	Uncertainty mA
1	40Ma-100A	2	2.002	-0.002	0.045

Resistance calibration

Sl.no	Range	Master readings ohms	Calibrated values ohms	Error observed ohms	Uncertainty Ohms
1	40	39	39.01	0.001	0.007

CALIBRATION FIELD: RPM Calibration

Sl.no	Range	Master readings RPM	Calibrated values RPM	Error observed RPM	Uncertainty RPM
1	1RPM to 4000RPM	1600RPM	1599RPM	1RPM	0.055


4.remarks

4.1 uncertainty estimated for a confidence level of 95% at the k factor=2

4.2the reported results are valid only for the condition of the inst./gauge received at the time of and under the condition of the calibration

4.3 DUC -device under calibration

Calibration label /status : Calibration result pass & fail are indicated


MEDICAL OFFICER
 GOVT. PRIMARY HEALTH CENTRE
 KOTHAMANGALAM
 PUDUKOTTAI DISTRICT

CALIBRATION ENGINEER		AUTHORIZED SIGN	This Report Refers Only To The Sample Submitted And May Not Be Reproduces in Full Or In Part, Without Written Permission From Roots Metrology & Testing Laboratory, Coimbatore, cell-9585840192
----------------------	---	-----------------	---