

				CALIBE	RATI	ON	CERT	IFIC/	ATE		
CERTI	FICATE NO	D:	SBS/CL/23	/11954			N	IECHAN	CAL	F	Page No:1 of 1
Issue Date				29-09-2023							
SRF No & Date					SRF/23/00424-0004 & 28-09-2023						
Receipt Date					28-09-2023						
Calibration Date						28-09-2023					
Calibra	ation Due					27-09-2024					
	ner Name &										
GOVE	RNMENT L	JPGRA	DED PRIMAR	RY HEALTH C	CENTRE						
PANA	PAKKAM-6	31052.							*		
				Detail	s of Dev	vice U	Inder Calibra	tion (DU	C)		
Descri	scription : CENTRIFUGE			Make & Model		:	NA & BDI-1	52			
Range	nge : 3500 RPM			Sr. No		:	NA				
Resolution : 10 RPM			Identification No : NA								
DUC (	Condition	:	Satisfactor			Loca		:	LABORAT	ORY	
			Envir	onmental Co	nditions	s & St	andard Oper	ating Pr	ocedure Details		
Environmental Details Temperature			Temperature	25.5°C Relative H		Relative Hur	midity 52%Rh				
Calibration Procedure No SBS/CP/ML				04 Calibrat		Calibration d	done at ONSITE				
				A	Refere	nce S	Standards De	tails			
S.No	No Description				Make/ SI No:		:	Certificate No		Validity	
1	Digital Tachometer			LINE SEIKI / 175-0034V		JRPM-CCTR-A&S-2023-0013		0013	09-06-2024		

## CALIBRATION RESULTS

S.No	DEVICE UNDER CALIBRATION	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY	
	RPM	RPM	RPM	%	
1	500	499.0	1,0	4.2	
2	1500	1499.6	0.4	4.2	
3	2500	2499.5	0.5	4.2	

## REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use. 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By (Calibration Engineer)



Authorised Sig	natory
(Quality Manager) (C.SIVABALAN)	(Chief Executive)