



| CALIBRATION REPORT | |
|---|--|
| Report No: HBMS/CAL/23-24 | Calibration Date: 27.09.2023 |
| | Calibration Valid: 26.03.2024 |
| 1.1. CUSTOMER DETAILS: | |
| Name and Address of the organization | M/s, GOVT. PRIMARY HEALTH CENTRE, ANAIMALAI BLOCK, V.KALIYAPURAM- 642129. |
| 1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT): | |
| EQUIPMENT NAME | PIPETTE |
| I Manufactured by | MERILETTE |
| II Model | 5µl - 50µl |
| III Serial No | 22205639 |
| IV Location | LABORATORY |

Calibration standards used :


| Master instrument | Model | Serial no | Date of calibration | Calibration due on |
|--------------------------|---------|------------|---------------------|--------------------|
| Electrical compact scale | SF-400C | SF400-2240 | 07-OCT-2022 | 06-OCT-2023 |

Details of the machine: machine calibrated and checked output.

| Number of measurement | Channel 1 (Volume) | | |
|-----------------------|--------------------|-------|-------|
| | 05µl | 20µl | 50µl |
| 1 | 05.05 | 20.08 | 50.97 |
| 2 | 05.06 | 20.07 | 20.02 |
| 3 | 05.01 | 20.05 | 20.96 |
| 4 | 05.97 | 20.99 | 20.02 |
| 5 | 05.01 | 20.08 | 20.09 |
| 6 | 05.02 | 20.95 | 20.04 |

For Heart Beat Medical Systems.

(Authorized signatory)



MD Reg No: TN/CBE/MD42/00016

99400 28355, 9751206106

No.42/1, Navaindia Road, Peelamedu,

GSTIN: 33AAJFH1833M1ZB

88833 15662

CBE - 04, hbmedicalsystem@gmail.com