



Trust & Care



heartbeat

Medical Systems

ELECTRICAL SAFETY TESTING/ CALIBRATION

Report No : HBMS/CAL/22-23

Calibration date : 14.12.2022

Calibration Valid : 13.06.2023

1.1 CUSTOMER DETAILS:

Name and Address of the organization

M/S. PRIMARY HEALTH CENTRE ,
NAMAGIRIPETTAL,NAMAKKAL.

1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):

EQUIPMENT NAME

SEMI AUTO ANALYZER

I Manufactured by

II Model

ROBONIC

iii Serial No

ATCD0660621RBIL

Iv Location

LABORATORY

2.CONDITION OF THE ITEM WHEN RECEIVED

No visible damages and in working order

3.TEST DONE

Electrical safety/ Performance Testing

4. STATUS

Manufactures Specification	User specification	Within specification	Out of specification	Calibration	Electrical Safety Test	Performance Analysis
√	-	√	-	-	√	-

No. 23, 2nd Floor, jayasimmapuram,
Pappanaikenpalayam, Coimbatore – 37.

99400 28355, 97512 06106
88833 15662

hbmedicalsystem@gmail.com
hbmssales@gmail.com



5. TRACEABILITY DETAILS OF INSTRUMENTS USED FOR TESTING

S.no	Name of the instrument	Make	Model	Serial no	Cal Due
1.	Electrical safety analyser	Fluke	Es612	3005050	07-OCT-2023

6. ELECTRICAL SAFETY TEST

S.NO	PARAMETER	OBSERVED VALUE	ACCEPTANCE LIMITS AS PER THE STD	REMARKS
1.	Protective earth resistance	0.129Ω	<0.3 Ω	Ok
2.	Earth leakage	134.21μA	<210 μA NC	Ok
3.	Patient leakage current	NA	<100 μA BF	Ok
4.	Equipment current	0.3A	As per the manufacture spec	Ok
5.	Mains voltage (P-E)	230.5V	-	Ok
6.	Mains voltage (P-N)	231.2V	-	Ok
7.	Mains voltage (N-E)	1.0V	<5V	Ok
8.	Enclosure leakage current	50 μA	<100 μA (normal)	Ok
9.	Insulation resistance	OL	>100 M Ω	Ok

S.NO	7.Remarks
1.	This report is applicable to the sample tested only
2.	The instrument used for testing are valid calibration and are traceable to national standards
3.	Parameter of the DUT were verified and found to be within the specified limits

Verified by :

[Signature]
MEDICAL OFFICER
 Govt. Primary Health Centre,
 Namagiripet - 641115,
 Naniakkal District.

Calibration Engineer

[Signature]
 HBSMS

No. 23, 2nd Floor, Jayasimmapuram,
 Pappanaikenpalayam, Coimbatore - 37.

99400 28355, 97512 06106
 88833 15662

hbmedicalsystem@gmail.com
 hbmsales@gmail.com