

## **CHRISTIAN MEDICAL COLLEGE**

## DEPARTMENT OF CLINICAL BIOCHEMISTRY

## CMC EXTERNAL QUALITY ASSURANCE SCHEME

## MONTHLY SUMMARY REPORT - JULY 2023



Lab Name

Govt. PHC, Sokkanur

Lab No

12041

Constituent Group

Chemistry III

Date of Result Entered:

19/07/2023

PT item

Lyophilized human serum based

Date of Report Published:

07/08/2023

SI.No	` Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your	CDI	
1	i					CV	SD	Value	SDI	U
1	GLUCOSE III	GOD-POD III	Transasia / Erba	24	109.83	6.83	7.50	110.4 mg/dL	0.08	3.06
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Transasia / Erba	18	5.79	3.00	0.17	5.7	-0.52	0.08
3	CHOLESTROL III	CHOD-PAP III	Transasia / Erba	19	112.64	4.38	4.93	mg/dL 114.2	0.32	2.26
								mg/dL		

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS:

Govt. PHC, Sokkanur THE MEDICAL OFFICER, SOKKANUR NALLATIPALAYAM Coimbatore **TAMILNADU** 

> **Coordinator Contact Details:** Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102

Panela Christudos

Dr. Pamela Christudoss **CMC EQAS Coordinator** Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed. Data in CMC EQAS reports is confidential CMC EQAS does not sub contract any components \*\*\*\*\*\* End of Report \*\*\*\*\*

சொக்கனூர், கிணத்துகடவு - 642 108.

Scanned with OKEN Scanner