

	8 8			CAL	BRA	TION	CERT	TIFICATE	<u> </u>		
CERTIFICATE NO: SBS/CL/23/1172			3/11725	11725		М	ECHANICAL	P	age No:1 of 1		
Issue Date				25-09-2023							
SRF No & Date						SRF/23/00400-0002 & 23-09-2023					
Receipt Date						23-09-2023					
Calibration Date						23-09-2023					
Calibration Due						22-09-2024					
Customer Name & Address											
GOVERNMENT PRIMARY HEALTH CENTRE,											
SOK	ANUR-642	109									
				De	tails of D	evice Un	der Calibra	ition (DUC)			
Description : CENTR		CENTRIF	ENTRIFUGE		Make & Model			B.D INSTRUMENTATION & BDI-152			
Range :			3000		Sr. No		:	95111159			
Resolution :			10	10 RPM		Identification No		:	NA		
DUC Condition : Satisfactory		у	Location		18 1 251 1 3	:	LABORATORY				
			Envi	ronmental	Condition	ns & Star	dard Oper	ating Procedure	e Details		
Environmental Details Temperature				e: 24.3 °	2: 24.3 °C R		ımidity	54 % Rh			
Calibration Procedure No SBS/CP/ML				/04	04 Calibration		done at ONSITE				
					Refer	rence Sta	ndards De	tails			
S.No	No Description			Make/ S	Make/ SI No:		Certificate No		Validity		
	Digital Tachometer			LINE SE	LINE SEIKI / 175-0034V		JRPM-CCTR-A&S-2023-0013		09-06-2024		

## **CALIBRATION RESULTS**

S.No	DEVICE UNDER CALIBRATION READINGS	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY (±)	
	RPM	RPM	RPM	%	
1	1000	999.8	0.2	4.2	
2	2000	1999.7	0.3	4.2	
3	3000	. 2999.9	0.1	4.2	

## REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

ば는 (Calibration Engineer) K.SATHYAMOORTHY



Quality Manager (C.SIVABALAN)

Chief Executive

**Authorised Signatory** 

## SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087, E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.

