

		7	CAL	IBRAT	ION	I CERT	TIFICA	<u>TE</u>	
CERTIFICATE NO: SBS/CL/23/11726				MEDICAL DEVICES			Page No:1 of 1		
Issue Date					25-09-2023				
SRF No & Date					SRF/23/00400-0003 & 23-09-2023				
Receip	t Date				23-09-2023				
Calibration Date				23-09-2023					
Calibra	tion Due				22-09-2024				
Custo	mer Name & Add	ress							
GOVE	RNMENT PRIMA	RY HEA	LTH CENTR	E,					
SOKK	ANUR-642109.								
				Details of De	vice U	nder Calibra	tion (DUC)		
Descrip	Description ELECTRICAL SAFET (MICROSCOPE)		Y	Make	& Model		MAGNUS & MLX		
Range		MULTI			Sr. N	0		NA	
Resolut	tion .	MULTI			Identi	fication No		NA	
DUC C	ondition :	SATIS	FACTORY		Local	ion		LABORATORY	
			Environ	nental Condi	tions	& Calibration	Procedure I	Details	
Environmental Details Temperature:2		25.6°C Relative Hun		nidity 52% RH					
Calibration Procedure No SBS/CP/MD/2		9 Calibration de		one at ONSITE					
				Refere	ence S	tandards De	tails		
S.No C	S.No Description M.		Make/ SI No:		Certificate No		Validity		
1 E	1 Electrical Safety Analyser		Rigel Medical & 44L-1059		M-230809-16-4		10-08-202		

ELECTRICAL	SAFETY
------------	--------

RESULTS						
S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)			
1	Insulation Resisitance	Measured values in MΩ	Uncertainty in % (±)			
	>20MΩ	97	13 92			
2	Earth Leakage	Measured values in µA	Uncertainty in % (±)			
	<5000µAfor B.BF.CF	125	8.0			
3	Enclosure Leakage	Measured values in μA	Uncertainty in % (±)			
	<500µAfor B,BF,CF	225	7.2			

REMARKS

- 1 This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory,
- 2 The user should determine the suitability of the instrument for its intended use.
- 3 The recalibration interval should be determined on the user requirement
- 4 The results stated in this certificate relate only to the item calibrated
- 5 The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6 Equipment used for Calibration were calibrated & traceable to National & International Standards

Calibrated By

Authorised Signatory

F3F4 (Calibration Engineer) K.SATHYAMOORTHY





Chief Executive

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087, E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.

