

CALIBRATION CERTIFICATE							
CERTIFICATE NO: SBS/CL/23/11949		MEDICAL DEVICES			Page N	lo:1 of 1	
Issue Date	29-0	9-2023					
SRF No & Date		SRF/23/00426-0003 & 28-09-2023					
Receipt Date		28-09-2023					
Calibration Date		28-09-2023					
Calibration Due	27-0	27-09-2024					
Customer Name & Address							
GOVERNMENT PRIMARY HEALTH CENTRE	Ξ,						
MELKALATHUR-632502, VELLORE DISTRIC	Т.						
	Details of Device L	Jnder Calibrati	ion (DUC)				
Description : SEMI AUTO ANALYZER		e & Model	:	VERITAS & SAB 100			
Range : MULTI	Sr. N		:	AT3730317SAB			
Resolution : MULTI	Ident	ification No	:	NA			
DUC Condition : SATISFACTORY		tion	:	LABORATORY			
Environmental Conditions & Calibration Procedure Details							
Environmental Details Temperature:25.6°C		Relative Humidity 52% RH					
Calibration Procedure No SBS/CP/MD/20		Calibration done at ONSITE					
Reference Standards Details							
S.No Description Make/ SI No:			Certificate No			Validity	
1 Electrical Safety Analyser Rigel Medical		HL-1059 M-230809-16-4		ľ	10-08-2024		

ELECTRICAL SAFETY

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±	
1	Insulation Resisitance	Measured values in MΩ	Uncertainty in % (±)	
	>20MQ	87	13.92	
2 Earth Leakage	Measured values in µA	Uncertainty in % (±)		
-	<5000µAfor B,BF,CF	195	9.3	
3	Enclosure Leakage	Measured values in µA	Uncertainty in % (±)	
	<500µAfor B,BF,CF	241	8.4	

REMARKS

1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.

2. The user should determine the suitability of the instrument for its intended use.

3. The recalibration interval should be determined on the user requirement.

4. The results stated in this certificate relate only to the item calibrated.

5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00. 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By 04. (Calibration Engineer) M.RAGUL

CAL	SOLUN
OME	101 10 N
18 000	0)0
NIHST	VNS

1.51 Technical Manager

Authorised Signatory

(C.SHANMUGARAJ)

Chief Executive

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087, E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.