

CALIBRATION CERTIFICATE

Certificate No: SBS/CI/23/11821

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE, SOKKANUR-642109.	SRF No.	SRF/23/00400-0024
	SRF Date	27-09-2023
	Date of Receipt	25-09-2023
	Date of Calibration	27-09-2023
	Due Date for Calibration	26-09-2024
	Issue Date	28-09-2023

Details of Unit Under Calibration

Description	MICRO PIPETTE	Make	THERMO SCIENTIFIC
Range	100-1000µl	Model	FINNPIPETTE F3
Resolution	1µl	Material	PVC
Serial Number	RW02864	Operating Range	100-1000µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition

Temperature (°C)	23.5	Humidity (%RH)	52	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

Calibration Method Used

Standard Used

Sl. No.	Description	ID.No. / Sl. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSP122/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319

Result of Calibration in µl


Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (t)
1	100	99.94	99.96	99.94	99.92	99.93	99.95	-0.05	0.02	0.47
		99.94	99.95	99.96	99.97	99.98				
2	500	499.85	499.86	499.87	499.89	499.86	499.87	-0.13	0.02	0.47
		499.86	499.87	499.89	499.85	499.89				
3	1000	999.96	999.97	999.98	999.96	999.95	999.96	-0.04	0.01	0.47
		999.98	999.94	999.96	999.95	999.95				

Remarks

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated
- Equipment used for Calibration were calibrated & traceable to National & International Standards
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- Calibration Liquid Used: Distilled or Deionized water containing Grade 3 as specified in ISO 3696.

Calibrated By,

M. Adhithyan
(Calibration Engineer)
MADHIBAN



Authorised by:

C. Sivabalan
(Quality Manager/Chief Executive)
C.SIVABALAN

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.