

			CALIBI	<u>RATIO</u>	N CERT	TIFICA	TE		
CERTIF	ICATE NO:	SBS/CL/23	3/11952			MECHANIC	CAL	Page No:	1 of 1
Issue D	ate			29-0	09-2023				
SRF No & Date			SRI	SRF/23/00425-0004 & 28-09-2023					
Receipt	Date			28-	28-09-2023				
Calibration Date			28-	28-09-2023					
Calibration Due			27-	27-09-2024					
Custom	er Name & Ad	dress							
GOVER	RNMENT CON	MUNITY HEAL	TH CENTRE,						
PUNNA	N-604405.								
			Detail	s of Device	Under Calibr	ation (DUC	;)		
Descrip	otion :	CENTRIF	JGE	Ма	ke & Model	:	REMI & C-8	354/4	
Range	:		3500 RPM	Sr.	No	:	ZGFN-1198	38	
Resolut	tion :		10 RPM	lde	ntification No	:	NA		
	ondition :	Satisfacto	γ	Loc	cation	:	LABORATO	DRY	
		Envir	onmental Co	nditions &	Standard Ope	rating Pro	cedure Details		
Enviror	Environmental Conditions & Standard Operating Procedure Details mental Details Temperature: 25.5°C Relative Humidity 52%Rh								
Calibration Procedure No SBS/0		SBS/CP/ML/	BS/CP/ML/04		Calibration done at				
				Reference	Standards D	etails			
S.No	No Description		Make/ SI N	0:	Certificate No			Validity	
1	1 Digital Tachometer			LINE SEIKI / 175-0034V		JRPM-CCTR-A&S-2023-0013		013	09-06-2024

CALIBRATION RESULTS

S.No	DEVICE UNDER CALIBRATION	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY	
	RPM	RPM	RPM	%	
1	1000	999.0	1.0	4.2	
2	2000	1999.7	0.3	4.2	
3	3000	2999.5	0.5	4.2	

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

(Calibration Engineer) M.RAGUL



Authorised Signatory

(Quality Manager) (C.SIVABALAN)

(Chief Executive)

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087, E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.