

CALIBRATION CERTIFICATE: PRETEST TOUCH			
Customer Name: The Medical Officer		Model : Prietest touch	
Address: Govt Primary Health Centre		S.No : ATP0241121RBK	
No:24 Veerapandi		Date of Calib., : 26.09.2023	
		Due date Calib., : 25.09.2024	
		Calibrated By : krishnakumar	
UPS Power Supply	Observed Reading	Remarks	
Earth To Neutral (0 - 3V AC)	0VAC		
Earth to Phase (225 - 230VAC)	226VAC	Measured Voltages are within the specifications.	
Phase to Neutral (225 - 230VAC)	225VAC		
Measuring Unit :			
	Expected Reading	Observed Reading	Remarks
Lamp Voltage	6 v	6.01v	No adjustments
	Good/ No Good	Remarks	
Pump	Good		
Filterwheel motor	Good	Performance of the parts are in	
Flowcell condition	OK	well within working specifications.	
Lens check	OK		
Filters Check:			
Wavelength	Range	Amplitude	Delta
340	3.00-11.00	3.921	0.001
405	3.00-11.00	5.214	0.004
505	3.00-11.00	5.487	0.001
545	3.00-11.00	5.216	0.001
578	3.00-11.00	5.458	0.003
630	3.00-11.00	5.224	0.001