



CALIBRATION REPORT	
Report No : HBMS/CAL/23-24	Calibration Date : 23.09.2023
	Calibration valid : 22.03.2024
<b>1.1.CUSTOMER DETAILS:</b>	
Name and Address of the organization	<b>M/s, GOVERNMENT PRIMARY HEALTH CENTRE ,24-VEERAPANDI, COIMBATORE.</b>
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>	
EQUIPMENT NAME	MICRO PIPETTE
i Manufactured by	THERMO SCIENTIFIC
ii Model	FINNE PIPETTE
iii Value	100µl-1000µl
iv Serial No	RW02857

**Calibration standards used :**

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

**Calibration test Report.**

Number of measurement	Channel 1 (Volume)		
	100µl	500µl	1000µl
1	100.05	500.11	1000.12
2	100.06	500.12	1000.03
3	100.02	500.01	1000.07
4	100.04	500.02	1000.08
5	100.00	500.00	1000.09
6	100.08	500.07	1000.02

**For Heart Beat Medical Systems,**

  
(Authorized signatory)





CALIBRATION REPORT	
Report No : HBMS/CAL/23-24	Calibration Date : 23.09.2023 Calibration valid : 22.03.2024
<b>1.1.CUSTOMER DETAILS:</b>	
Name and Address of the organization	<b>M/s, GOVERNMENT PRIMARY HEALTH CENTRE ,24-VEERAPANDI, COIMBATORE.</b>
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>	
EQUIPMENT NAME	MICRO PIPETTE
i Manufactured by	VISION PLUS
ii Model	
iii Value	100µl-1000µl
iv Serial No	UPHC/24V/MP/001

**Calibration standards used :**

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

**Calibration test Report.**

Number of measurement	Channel 1 (Volume)		
	100µl	500µl	1000µl
1	100.04	500.11	1000.13
2	100.01	500.11	1000.31
3	100.05	500.13	1000.32
4	100.13	500.12	1000.12
5	100.12	500.31	1000.12
6	100.12	500.12	1000.12

For Heart Beat Medical Systems,

  
(Authorized signatory) 



CALIBRATION REPORT	
Report No : HBMS/CAL/23-24	Calibration Date : 23.09.2023
	Calibration valid : 22.03.2024
<b>1.1.CUSTOMER DETAILS:</b>	
Name and Address of the organization	M/s, GOVERNMENT PRIMARY HEALTH CENTRE ,24-VEERAPANDI, COIMBATORE.
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>	
EQUIPMENT NAME	MICRO PIPETTE
i Manufactured by	THERMOSCIENTIFIC
ii Model	FINNIPETTE
iii Value	10 µl -100 µl
iv Serial No	RW01623

**Calibration standards used :**

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

**Calibration test Report.**

Number of measurement	Channel 1 (Volume)		
	10µl	50µl	100µl
1	10.03	50.06	100.14
2	10.04	50.09	100.12
3	10.05	50.02	100.10
4	10.09	50.08	100.08
5	10.00	49.94	100.09
6	10.08	50.07	100.12

For Heart Beat Medical Systems,

(Authorized signatory)

