




# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - AUGUST 2023



PC-1024

Lab Name Govt. PHC, 24 Veerapandi

Lab No 12053

Constituent Group Chemistry III

Date of Result Entered : 16/08/2023

PT item Lyophilized human serum based

Date of Report Published : 01/09/2023

| Sl.No | Analyte        | Method / Principle Name               | Analyzer Name                                | No of Participants | DV     | Participants |      | Your Value  | SDI   | U    |
|-------|----------------|---------------------------------------|--|--------------------|--------|--------------|------|-------------|-------|------|
|       |                |                                       |  |                    |        | CV           | SD   |             |       |      |
| 1     | GLUCOSE III    | GOD-POD III                           | Robonik ( Semi automatic / Fully Automatic ) | 1841               | 175.58 | 5.13         | 9.00 | 172.6 mg/dL | -0.33 | 0.42 |
| 2     | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik ( Semi automatic / Fully Automatic ) | 1852               | 1.76   | 6.77         | 0.12 | 1.69 mg/dL  | -0.59 | 0.01 |
| 3     | CHOLESTROL III | CHOD-PAP III                          | Robonik ( Semi automatic / Fully Automatic ) | 1764               | 113.80 | 4.07         | 4.64 | 122.7 mg/dL | 1.92  | 0.22 |

| SDI Range                       | Interpretation                           |
|---------------------------------|--|
| Within -1.00 to +1.00           | Excellent.                               |
| Within $\pm 1.01$ to $\pm 2.00$ | Good.                                    |
| Within $\pm 2.01$ to $\pm 2.99$ | Accept with caution. Warning Signal.     |
| Beyond $\pm 3.0$                | Unacceptable performance. Action Signal. |

LAB ADDRESS :

Govt. PHC, 24 Veerapandi  
THE MEDICAL OFFICER, 24 VEERAPANDI  
Coimbatore  
TAMILNADU641018

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

2023.9.29 09:45

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

\*\*\*\*\* End of Report \*\*\*\*\*