

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Kombukarananthal	Details About Robust Analysis
Lab No	12649	Detail About Monthly Summary
Month	March	Detail about SDI
Year	2023	
Constituent Group	Chemistry III	

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 19/03/2023


Date of Report Published : 04/04/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1875	91.48	7.46	6.62	84.8 mg/dL	-0.98	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08	0.12	1.21 mg/dL	-0.99	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1761	110.93	5.23	5.80	103.2 mg/dL	-1.33	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out


 MEDICAL OFFICER
 GOVT. PRIMARY HEALTH CENT
 KOMBUKARANANTHAL, SIVADANGAL L.

VIEW LAB MONTHLY SUMMARY


Lab Name	Govt. PHC, Kombukaranenthal	Details About Robust Analysis	
Lab No	12649	Detail About Monthly Summary	
Month	April	Detail about SDI	
Year	2023		
Constituent Group	Chemistry III		
Click on the analyte to view Graphical Data		All Analyser Result	Print
		Date of Result Entered : 20/04/2023	
		Date of Report Published : 06/05/2023	

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SO	Your Value	SDI	LI
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	124 mg/dL	0.63	0.34
2	CREATININE III	Jaffee Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.1 mg/dL	-0.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	117 mg/dL	0.93	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out


 MEDICAL OFFICER
 GOVT. PRIMARY HEALTH CENTRE
 KOMBUKARANENTHAL, SIVAGANGAI, DI

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Kombukaranenthal
 Lab No 12649
 Month May
 Year 2023
 Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 20/05/2023

Date of Report Published : 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	192 mg/dL	1.18	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.78 mg/dL	0.15	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	121 mg/dL	1.10	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out


 MEDICAL OFFICER
 GOVT. PRIMARY HEALTH CENTRE
 KOMBUKARANENTHAL, SVAGANGAI DT

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Kombukaranenthal
 Lab No 12649
 Month June
 Year 2023
 Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 20/06/2023

Date of Report Published : 04/07/2023


Sl.No	Analyte	Method / Principle Name	Analyser Name	No. of Participants	DV	Participants		Year Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	279 mg/dL	-0.51	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.5 mg/dL	-0.41	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	106 mg/dL	-1.52	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

M. S. S. S.


 MEDICAL OFFICER
 GOVT. PRIMARY HEALTH CENT.
 KOMBUKARANENTRAL, SIVAGANGAI D.

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Kombukaranenthal	Details About Robust Analysis
Lab No	12649	Detail About Monthly Summary
Month	July	Detail about SDI
Year	2023	
Constituent Group	Chemistry III	

Click on the analyte to view Graphical Data

Date of Result Entered : 19/07/2023

Date of Report Published : 07/08/2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SO			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	101 mg/dL	-1.38	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	5.7 mg/dL	-0.37	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	112 mg/dL	-0.21	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

M.P.R
Siva Anand

[Signature]
MEDICAL OFFICER
GOVT. PRIMARY HEALTH CENTRE
KOMBUKARANENTHAL, SIVAGANGAI

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Kombakaranenthal
 Lab No 12649
 Month August
 Year 2023
 Constituent Group Chemistry III

Details About Robust Analysis
 Detail About Monthly Summary
 Detail about SDI

Click on the analyte to view Graphical Data

Date of Result Entered : 19/09/2023

Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	OV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	168.3 mg/dL	-0.81	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.76 mg/dL	0.00	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1784	113.80	4.07	4.64	112.4 mg/dL	-0.30	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Monthly Address	Modify Methods	Small Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

M. P. D.
 D.A. Nishitha

மருத்துவ அலுவலர்
 அரசு ஆரம்ப சுகாதார நிலையம்
 கொம்புகாரனேந்தல்
 தருவங்காடு மாவட்டம்