

# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
 CMC EXTERNAL QUALITY ASSURANCE SCHEME  
 MONTHLY SUMMARY REPORT - JULY 2023



Lab No 10872  
 Date of Result Entered : 20/07/2023  
 Date of Report Published : 07/08/2023

Govt. UPHC, Mannargudi  
 Chemistry III  
 Lyophilized human serum based

Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
					CV	SD			
LUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1876	109.35	5.55	6.07	118 mg/dL	1.43	0.28
UREA III	Urease UV / GLDH III	Robonik ( Semi automatic / Fully Automatic )	184	51.41	11.14	5.73	61 mg/dL	1.67	0.84
CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1850	5.83	6.00	0.35	6.7 mg/dL	2.49	0.02
CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1776	113.00	4.12	4.66	119 mg/dL	1.29	0.22

SDI Range	Interpretation
-1.00 to +1.00	Excellent.
-1.01 to ±2.00	Good.
-1.01 to ±2.99	Accept with caution. Warning Signal.
±3.0	Unacceptable performance. Action Signal.

ADDRESS :  
 Govt. UPHC, Mannargudi  
 ANI PALAYAM ROAD, MANNARGUDI  
 Mannargudi  
 DU614001

Coordinator Contact Details:  
 Email: clinqc@cmcvellore.ac.in  
 Contact Number: 0416-2283102

*Pamela Christudoss*  
 Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential  
 CMC EQAS does not sub contract any components  
 \*\*\*\*\* End of Report \*\*\*\*\*

சென்னை சி.மீ.கல்.வெ.லூர்  
 சி.மீ.கல்.வெ.லூர் - 014 001

**CHRISTIAN MEDICAL COLLEGE**

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JUNE 2023



PC-1024

Lab Name: Govt. UPHC, Mannargudi      Lab No: 10872  
 Constituent Group: Chemistry III      Date of Result Entered: 16/06/2023  
 Test Item: Lyophilized human serum based      Date of Report Published: 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1823	286.63	5.17	14.82	265 mg/dL	-1.46	0.69
2	UREA III	Urease UV / GLDH III	Robonik ( Semi automatic / Fully Automatic )	192	32.69	12.96	4.24	33 mg/dL	0.07	0.61
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1848	2.57	6.64	0.17	2.3 mg/dL	-1.58	0.01
4	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1804	113.73	4.46	5.07	115 mg/dL	0.25	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

## LAB ADDRESS :

Govt. UPHC, Mannargudi  
 RUKKUMANI PALAYAM ROAD, MANNARGUDI  
 Mannargudi  
 TAMILNADU614001

நகர ஆரம்ப சுகாதார நிலையம்  
 மன்னார்குடி - 614 001



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - MAY 2023



PC-1024

Lab Name	Govt. UPHC, Mannargudi	Lab No	10872
Constituent Group	Chemistry III	Date of Result Entered :	18/05/2023
PT Item	Lyophilized human serum based	Date of Report Published :	05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	182 mg/dL	0.12	0.44
2	UREA III	Urease UV / GLDH III	Robonik ( Semi automatic / Fully Automatic )	163	27.91	13.39	3.74	33 mg/dL	1.36	0.59
3	CREATININE III	Jaffes Kinetlc - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1854	1.76	7.48	0.13	1.7 mg/dL	-0.45	0.01
4	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1798	114.60	5.08	5.83	115 mg/dL	0.07	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

## LAB ADDRESS :

Govt. UPHC, Mannargudi  
RUKKUMANI PALAYAM ROAD, MANNARGUDI  
Mannargudi  
TAMILNADU614001

Coordinator Contact Details:  
Email:clinqc@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*

பல அருள் சான்ற சீலையம்  
மேற்கூறிய - 614 001

8/18/23, 9:19 AM

External Quality Assurance Scheme - Print Monthly Summary



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - JANUARY 2023



PC-1024

Lab Name Govt. UPHC, Mannargudi  
Constituent Group Chemistry III  
PT item Lyophilized human serum based

Lab No  
Date of Result Entered :  
Date of Report Published :

10872  
20/01/2023  
09/02/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1890	119.87	6.38	7.65	114 mg/dL	-0.77	0.35
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1865	2.21	10.02	0.22	2.3 mg/dL	0.41	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1794	112.14	6.03	6.76	106 mg/dL	-0.91	0.32

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

## LAB ADDRESS :

Govt. UPHC, Mannargudi  
RUKKUMANI PALAYAM ROAD, MANNARGUDI  
Mannargudi  
TAMILNADU614001

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*

சென்னை - 614 001



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - FEBRUARY 2023



Lab Name: Govt. UPHC, Mannargudi  
Constituent Group: Chemistry III  
PT Item: Lyophilized human serum based

Lab No: 10872  
Date of Result Entered: 24/02/2023  
Date of Report Published: 08/03/2023

SLNo	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1868	155.11	6.56	10.18	159 mg/dL	0.38	0.47
2	UREA III	Urease UV / GLDH III	Robonik ( Semi automatic / Fully Automatic )	175	30.39	12.93	3.93	35 mg/dL	1.17	0.59
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1872	3.40	9.00	0.31	3 mg/dL	-1.31	0.01
4	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1788	112.79	5.23	5.90	105 mg/dL	-1.32	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

## LAB ADDRESS :

Govt. UPHC, Mannargudi  
RUKKUMANI PALAYAM ROAD, MANNARGUDI  
Mannargudi  
TAMILNADU 614001

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

\*\*\*\*\* End of Report \*\*\*\*\*

நகர ஆரம்ப சுகாதார நிலையம்  
மன்னார்குடி - 614 001

**CHRISTIAN MEDICAL COLLEGE**

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - APRIL 2023



PC-1024

Govt. UPHC, Mannargudi

Lab No

10872

Chemistry III

Date of Result Entered :

19/04/2023

Lyophilized human serum based

Date of Report Published :

06/05/2023

Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
					CV	SD			
GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1832	119.45	6.04	7.21	118 mg/dL	-0.20	0.34
UREA III	Urease UV / GLDH III	Robonik ( Semi automatic / Fully Automatic )	169	42.32	10.68	4.52	46 mg/dL	0.81	0.70
CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1820	2.16	8.09	0.18	2 mg/dL	-0.91	0.01
CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1750	111.61	5.19	5.79	115 mg/dL	0.59	0.28

SDI Range	Interpretation
00 to +1.00	Excellent.
01 to $\pm$ 2.00	Good.
01 to $\pm$ 2.99	Accept with caution. Warning Signal.
3.0	Unacceptable performance. Action Signal.

ADDRESS :

C, Mannargudi

NI PALAYAM ROAD, MANNARGUDI

Vellore

Tamil Nadu  
614001

Coordinator Contact Details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss

CMC EQAS Coordinator

Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

\*\*\*\*\* End of Report \*\*\*\*\*

தலைவர் அலுவலகம்  
மன்னார்குடி - 614 001