



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JULY 2023



Lab Name	Govt. PHC, Vadaponparappi	Lab No	12274
Constituent Group	Chemistry III	Date of Result Entered	19/07/2023
PT Item	Lyophilized human serum based	Date of Report Published	07/08/2023

Sl. No.	Parameter	Method	Instrument/Reagent	Ref. Range	Result	CV (%)	CV (%)	CV (%)	CV (%)	CV (%)
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	435.4	0.28	
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	3.75	6.4	0.02	
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	5.72	4.55	115.3	0.22	

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. PHC, Vadaponparappi
Vadaponparappi
TAMIL NADU

Coordinator Contact Details:
Email: clinqs@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela C. ...
Dr. Pamela C. ...
CMC EQAS ...
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

dy
**MEDICAL OFFICER
PRIMARY HEALTH CENTRE
VADAPONPARAPPI-605 70**