



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
 CMC EXTERNAL QUALITY ASSURANCE SCHEME  
 MONTHLY SUMMARY REPORT - MAY 2023



Lab Name: Govt. PHC, Vadaponparappi      Lab No: 2023  
 Constituent Group: Chemistry III      Date of Result Entered: 20/05/23  
 PT Item: Lyophilized human serum based      Date of Report Published: 22/05/23

Sl. No.	Analyte	Method / Principle Name	Analyzer Name	Lot #	1	2	3	4	5	6
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	95.6 mg/dL	1.05	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1854	1.76	7.48	0.13	1.0 mg/dL	1.97	0.97
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1798	114.60	5.08	5.89	114.9 mg/dL	6.65	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :  
 Govt. PHC, Vadaponparappi  
 Vadaponparappi  
 TAMILNADU

Coordinator Contact Details:  
 Email: clinqc@cmcvellore.ac.in  
 Contact Number: 0416-2283102

*Pamela Christudoss*  
 Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential  
 CMC EQAS does not sub contract any components  
 \*\*\*\*\* End of Report \*\*\*\*\*

*S. Shanmugam*  
 MEDICAL OFFICER  
 PRIMARY HEALTH CENTRE  
 VADAPONPARAPPI-605 707