



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - MAY 2023



PC-1024

Lab Name Govt. UPHC, Bharathiyar Park Lab No 10726  
 Constituent Group Chemistry III Date of Result Entered : 18/05/2023  
 PT item Lyophilized human serum based Date of Report Published : 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Any Analyser (Automation / Semi Automation)	69	181.20	5.81	10.52	180.4 mg/dL	-0.08	2.53
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation)	68	1.75	7.05	0.12	1.68 mg/dL	-0.57	0.03
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	154	114.84	5.08	5.83	118.7 mg/dL	0.66	0.94

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

## LAB ADDRESS :

Govt. UPHC, Bharathiyar Park  
 MEDICAL OFFICER, BHARATHIYAR PARK, BHARATHI NAGAR, PARAMAKUDI  
 Paramakudi  
 TAMILNADU623707

*Pamela Christudoss*

Coordinator Contact Details:  
 Email: clinqc@cmcvellore.ac.in  
 Contact Number: 0416-2283102

Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

\*\*\*\*\* End of Report \*\*\*\*\*

*[Signature]*  
 MEDICAL OFFICER  
 Urban Health Centre  
 Bharathiyar Park  
 Paramakudi-623707