



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - MARCH 2023



PC-1024

Lab Name Govt. UPHC, Bharathiyar Park Lab No 10726  
Constituent Group Chemistry III Date of Result Entered : 18/03/2023  
PT item Lyophilized human serum based Date of Report Published : 04/04/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Any Analyser (Automation / Semi Automation)	74	92.52	7.71	7.13	94.35 mg/dL	0.26	1.66
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation)	70	1.32	7.52	0.10	1.41 mg/dL	0.91	0.02
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	148	111.55	4.29	4.78	110.66 mg/dL	-0.19	0.79

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

## LAB ADDRESS :

Govt. UPHC, Bharathiyar Park  
MEDICAL OFFICER, BHARATHIYAR PARK, BHARATHI NAGAR, PARMAKUDI  
Paramakudi  
TAMILNADU623707

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*

*Pamela Christudoss*  
**MEDICAL OFFICER**  
Urban Health Centre  
Bharathiyar Park  
Paramakudi-623707