

				CAL	IBRAT	101	V CER	TIFICA	ATE	
CER	TIFICATE N	O: SB	S/CL/23/	11855			N	MEDICAL D	EVICES	Page No:1 of 1
Issue Date						29-09-2023			1 490 11011 01 1	
SRF No & Date						SRF/23/00346-0001 & 27-09-2023				
Receipt Date						27-09-2023				
Calibration Date						27-09-2023				
Calibration Due						26-09-2024				
Custo	omer Name 8	Addre	ess							
GOV	ERNMENT	JRBAN	HEALT	H CENTRE,						
BHA	RATHIYAR F	PARK,	PARAMA	KUDI-62370	7.					
					Details of De	vice L	Inder Calibra	tion (DUC)		
_									ALPHA TECHNOLO	OGIES & ALPHA CHEN
Description : SEMI AUTO ANALY			UTO ANALYZ	ZER	Make	& Model	:	100		
Range : MULTI				Sr. N	0	:	BCAA17062651			
Resolution : MULTI				Identi	fication No	:	NA			
DUC Condition : SATISFACTORY				Locat		:	LAB			
					mental Condi	tions	& Calibration	Procedure	Details	
Environmental Details Temperature:25.6°C						Relative Humidity 52% RH		52% RH		
Calibration Procedure No SBS/CP/MD/20				Cambration done at		ONSITE				
					Refere	nce S	tandards De	tails		
S.No	No Description			Make/ SI No:		Certificate No		Validity		
1	Electrical Safety Analyser			Rigel Medical & 44L-1059		M-230809-16-4		10-08-202		

## **ELECTRICAL SAFETY**

## **RESULTS**


S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resisitance	Measured values in MΩ	Uncertainty in % (±)
	>20MΩ	87	13.92
2	Earth Leakage	Measured values in µA	Uncertainty in % (±)
	<5000µAfor B,BF,CF	195	0.3
3	Enclosure Leakage	Measured values in µA	Uncertainty in % (±)
	<500µAfor B,BF,CF	241	8.4

## REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- Equipment used for Calibration were calibrated & traceable to National & International Standards.
  Calibrated By

**M** •

Authorised Signatory





Techni	್ರ cal Manager
(C.SHA	ANMUGARAJ)

Chief Executive

## **SUNSHINE BIOMEDICAL SOLUTIONS**