

		CALIE	BRATI	10	N CER	TIFI	CAT	ΓΕ		
CERTIFICATE NO: SBS/CL/23/11856					MECH	IANICA		Page No:1 of 1		
Issue Date	•		2	29-0	9-2023					<u></u>
SRF No & Date			5	SRF	/23/00346-00	02 & 2	7-09-20)23		
Receipt Date					9-2023					
Calibration Date			2	27-0	9-2023					
Calibration Due			2	26-0	9-2024					
Customer Name & A	ddress									
GOVERNMENT URI	BAN HEALTH C	ENTRE,								
BHARATHIYAR PAR	RK,PARAMAKUE	01-623707.								
·		Deta	ils of Dev	ice	Under Calibr	ation	(DUC)			
Description	CENTRIF	UGE	N	Иак	e & Model		:	REMI & C-852		
Range		3500 RPM	s	Sr. N	0		:	ZCEN-13691		
Resolution	:	1 RPM	10	dent	ification No		:	NA		
DUC Condition	Satisfacto	ry	L	oca	tion		:	LAB		
	Envi	ronmental Co	onditions	& S	tandard Ope	rating	Proced	dure Details		
		Temperature	e: 25.5°C F		Relative Humidity		52%Rh			
Calibration Procedure No SBS/CP/M		SBS/CP/ML/	04 Calibi		Calibration d	oration done at		ONSITE		
			Referen	ice S	Standards De	etails				
S.No Description		Make/ SI No:			Certificate No				Validity	
1 Digital Tachometer		LINE SEIKI / 175-0034V		JRPM-CCTR-A&S-2023-0013				09-06-2024		

CALIBRATION RESULTS

S.No	DEVICE UNDER CALIBRATION	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY		
	RPM	RPM	RPM	%		
1	1000	998.9	1.1	4.2		
2	2000	1999.6	0.4	4.2		
3	3000	2999.6	0.4	4.2		

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

(Calibration Engineer)

M.RAGUL



Authorised Signatory

(Quality Manager) (Chief Executive)

(C.SIVABALAN)

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087, E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.