

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - FEBRUARY 2023



Lab Name

Govt. Block PHC, Palayampudur

Lab No

7351

Constituent Group

Chemistry III

Date of Result Entered :

21/02/2023

PT item

Lyophilized human serum based

Date of Report Published:

08/03/2023

SI.No	Analyte	Method / Principle Name	Analyzer Name	No of	DV	Partic	ipants	Your	1.46	1.46 0.47
		ivanje		Participants		C۷	SD	Value	SDI	
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1868	155,11	6.56	10.18	170 mg/dL	1.46	0.47
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1872	3.40	9.00	0.31	3 mg/dL	-1.31	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1788	112.79	5.23	5.90	105 mg/dL	-1.32	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. Block PHC, Palayampudur
THE MEDICAL OFFICER, PALAYAMPUR PHC, NALLAMPALLI, DHARMAPURI DIST.
Palayampudur
TAMILNADU636807

Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102 Panela Christudoss
Dr. Pamela Christudoss

CMC EQAS Coordinator
Christian Medical College, Vellore



#### **DEPARTMENT OF CLINICAL BIOCHEMISTRY**

#### CMC EXTERNAL QUALITY ASSURANCE SCHEME

#### **MONTHLY SUMMARY REPORT - JANUARY 2023**



Lab Name

Govt. Block PHC, Palayampudur

Lab No

7351

Constituent Group

Chemistry III

Date of Result Entered:

20/01/2023

PT item

Lyophilized human serum based

Date of Report Published:

09/02/2023

SI 11.	1000110	Method / Principle	* 1	No of	DV	Partici	pants	Your	0.02 0.3	
SI.No	Analyte	Name	Analyzer Name	Participants	υv	cv	SD	Value	SDI	0.02 0.35
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1890	119.87	6.38	7.65	120 mg/dL	0.02	0.35
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1865	2.21	10.02	0.22	1.9 mg/dL	-1.40	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1794	112.14	6.03	6.76	104 mg/dL	-1.20	0.32

SDI Range	Interpretation			
Within -1.00 to +1.00	Excellent.			
Within ±1.01 to ±2.00	Good.			
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.			
Beyond ±3.0	Unacceptable performance. Action Signal.			

LAB ADDRESS:

Govt. Block PHC, Palayampudur

THE MEDICAL OFFICER, PALAYAMPUR PHC, NALLAMPALLI, DHARMAPURI DIST.

Palayampudur

TAMILNADU636807

Panela Christudos

Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102 Dr. Pamela Christudoss CMC EQAS Coordinator Christian Medical College, Vellore



DEPARTMENT OF CLINICAL BIOCHEMISTRY

### CMC EXTERNAL QUALITY ASSURANCE SCHEME

### MONTHLY SUMMARY REPORT - MAY 2023



Lab Name

Govt. Block PHC, Palayampudur

Lab No

7351

Constituent Group

Chemistry III

Date of Result Entered:

19/05/2023

PT item

Lyophilized human serum based

Date of Report Published :

05/06/2023

SI.No	Analyte	Method / Principle	Analyzer Name	No of	DV	Partic	ipants	Your		
		Name		Participants	, o	CV	SD	Value	SDI	
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	182 mg/dL	0.12	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1854	1.76	7.48	0.13	1.72 mg/dL	-0.30	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1798	114.60	5.08	5.83	107 mg/dL	-1.30	0.27

SDI Range	Interpretation			
Within -1.00 to +1.00	Excellent.			
Within ±1.01 to ±2.00	Good.			
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.			
Beyond ±3.0	Unacceptable performance. Action Signal.			

LAB ADDRESS :

Govt. Block PHC, Palayampudur
THE MEDICAL OFFICER, PALAYAMPUR PHC, NALLAMPALLI, DHARMAPURI DIST.
Palayampudur
TAMILNADU636807

Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102 Panela Christudoss
Dr. Pamela Christudoss

CMC EQAS Coordinator Christian Medical College, Vellore



DEPARTMENT OF CLINICAL BIOCHEMISTRY

#### CMC EXTERNAL QUALITY ASSURANCE SCHEME

### MONTHLY SUMMARY REPORT - JUNE 2023



PC-1024

Lab Name

Govt. Block PHC, Palayampudur

Lab No

7351

Constituent Group

Chemistry III

Date of Result Entered :

19/06/2023

PT item

Lyophilized human serum based

Date of Report Published :

04/07/2023

SI.No	Analyte	Method / Principle	Analyzer Name	No of	DV	Partic	ipants	Your	0.56 -0.41	0.56 0.69
		Name	-mony-convenie	Participants	UV	CV	SD	Value	SDI	
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1823	286.63	5.17	14.82	295 mg/dL	0.56	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1848	2.57	6.64	0.17	2.5 mg/dL	-0.41	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1804	113.73	4.46	5.07	117 mg/dL	0.64	0.24

SDI Range		Interpretation				
Within -1.00 to +1.00 »		Excellent.				
Within ±1.01 to ±2.00		Good.				
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.				
Beyond ±3.0		Unacceptable performance. Action Signal.				

LAB ADDRESS :

Govt. Block PHC, Palayampudur

THE MEDICAL OFFICER, PALAYAMPUR PHC, NALLAMPALLI, DHARMAPURI DIST.

Palayampudur

TAMILNADU636807

Panela Christudoss

Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore



#### **DEPARTMENT OF CLINICAL BIOCHEMISTRY**

## CMC EXTERNAL QUALITY ASSURANCE SCHEME





Lab Name

Govt. Block PHC, Palayampudur

Lab No

7351

Constituent Group

Chemistry III

Date of Result Entered:

19/08/2023

PT item

Lyophilized human serum based

Date-of Report Published:

01/09/2023

SI.No	Analyte	Method / Principle	Analyzer Name	No of	DV	Partic	ipants	Your	20.	
	7	Name	Allalyzer Name	Participants	UV	CV	SD	Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	179 mg/dL	0.38	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	1.6 mg/dL	-1.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	108 mg/dL	-1.25	0.22

SDI Range	Interpretation			
Within -1.00 to +1.00	Excellent.			
Within ±1.01 to ±2.00	Good.			
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.			
Beyond ±3.0	Unacceptable performance. Action Signal.			

LAB ADDRESS :

Govt. Block PHC, Palayampudur

THE MEDICAL OFFICER, PALAYAMPUR PHC, NALLAMPALLI, DHARMAPURI DIST.

Palayampudur

TAMILNADU636807

Panela Christudoss

Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102 Dr. Pamela Christudoss CMC EQAS Coordinator Christian Medical College, Vellore