

VIEW LAB MONTHLY SUMMARY

Lab Name: **Govt. UPHC, Sandhaipettai**  
 Lab No: 10845  
 Month: **June**  
 Year: 2023  
 Constituent Group: Chemistry III

[Details About Robust Analysis](#)  
[Detail About Monthly Summary](#)  
[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 20/06/2023

Date of Report Published : 04/07/2023

SLNo	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Year Value	SD
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	297.3 mg/dL	0.72
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic)	1848	2.57	6.54	0.17	2.6 mg/dL	0.18
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	117.4 mg/dL	0.74

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

**MEDICAL OFFICER**  
 JESAN PRIMARY HEALTH CENTRE  
 SANDHAIPETTAI  
 QUARANTINE - 658 701

classmate