



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - MARCH 2023



12560  
19/03/2023  
04/04/2023

Lab No  
Date of Result Entered :  
Date of Report Published :

Lab Name Govt. PHC, Natham  
Constituent Group Chemistry III  
PT Item Lyophilized human serum based

SL No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	CV	SD	Your Value	SD	U
1	GLUCOSE III	GOD-POD III	Any Analyser (Automation / Semi Automation)	74	92.52	7.71	59 mg/dL	0.91	1.66
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation)	70	1.32	7.52	1.4 mg/dL	0.81	0.02
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	148	111.55	4.29	100 mg/dL	-2.41	0.79

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :  
Govt. PHC, Natham  
Natham  
TAMILNADU

Coordinator Contact Details:  
Email: [clinq@cmcvallore.ac.in](mailto:clinq@cmcvallore.ac.in)  
Contact Number: 0416-2283102

*Pamela Christoboss*

Dr. Pamela Christoboss  
CMC EQAS Coordinator  
Christian Medical College, Vellore