

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. PHC, Natham  
 Lab No: 12560  
 Month: April  
 Year: 2023  
 Constituent Group: Chemistry III

Click on the analyte to view Graphical Data

Details About Request Analysis  
 Detail About Monthly Summary  
 Detail about S01

Date of Result Entered : 19/04/2023

Date of Report Published : 09/05/2023

Sl.No	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	CV	SD	SD%	U	
1	GLUCOSE III	GOD-POD III	Any Analyser (Automation / Semi Automation )	74	118.33	5.77	6.83	116 mg/dL	-0.34	1.59
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation )	65	2.14	7.02	0.15	2.4 mg/dL	1.73	0.04
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation )	143	111.22	4.31	4.79	109 mg/dL	-0.46	0.80
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

Links

*Verified By  
 18/05/23  
 Dist. Microbiology*