



**CHRISTIAN MEDICAL COLLEGE**  
 DEPARTMENT OF CLINICAL BIOCHEMISTRY  
 CMC EXTERNAL QUALITY ASSURANCE SCHEME  
 MONTHLY SUMMARY REPORT - MAY 2023



Lab Name: Govt. PHC, Natham  
 Chemistry III  
 Consultant Group: Lyophilized human serum based  
 PT Item:

Lab No: 12960  
 Date of Result Entered: 15/05/2023  
 Date of Report Published: 05/06/2023

SL.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOO-POD III	Any Analyser (Automation / Semi Automation)	69	181.20	5.81	10.52	182 mg/dL	0.08	2.53
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation)	68	1.75	7.05	0.12	1.9 mg/dL	1.22	0.03
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	154	114.84	5.08	5.83	103 mg/dL	-2.03	0.94
		<b>SDI Range</b>	<b>Interpretation</b>							
		Within -1.00 to +1.00	Excellent.							
		Within ±1.01 to ±2.00	Good.							
		Within ±2.01 to ±2.99	Accept with caution. Warning Signal.							
		Beyond ±3.0	Unacceptable performance. Action Signal.							

LAB ADDRESS :  
 Govt. PHC, Natham  
 Natham  
 TAMILNADU

Coordinator Contact Details:  
 Email: cllnqc@cmcvellore.ac.in  
 Contact Number: 0416-2283102

*Pamela Christudass*  
 Dr. Pamela Christudass  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential