



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
 CMC EXTERNAL QUALITY ASSURANCE SCHEME  
 MONTHLY SUMMARY REPORT - JULY 2023



Lab Name: Govt. PHC, Natham

Constituent Group: Chemistry III

PT Item: Lyophilized human serum based

Lab No

Date of Result Entered :

Date of Report Published :

12560

19/07/2023

07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Any Analyser (Automation / Semi Automation)	75	109.71	5.43	5.95	107 mg/dL	-0.46	1.37
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Any Analyser (Automation / Semi Automation)	66	5.89	4.60	0.27	109 mg/dL	381.89	0.07
3	CHOLESTROL III	CHOD-PAP III	Any Analyser (Automation / Semi Automation)	148	112.78	4.33	4.88	6.6 mg/dL	-21.76	0.80
<b>SDI Range</b>										
				Excellent.						
Within -1.00 to +1.00				Good.						
Within ±1.01 to ±2.00				Accept with caution. Warning Signal.						
Within ±2.01 to ±2.99				Unacceptable performance. Action Signal.						
Beyond ±3.0										

LAB ADDRESS :  
 Govt. PHC, Natham  
 Natham  
 TAMILNADU

Coordinator Contact Details:  
 Email: dclngc@cmcvellore.ac.in  
 Contact Number: 0416-2283102

*Paula Christodoss*  
 Dr. Pamela Christodoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential