



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
 CMC EXTERNAL QUALITY ASSURANCE SCHEME  
 MONTHLY SUMMARY REPORT - MARCH 2023




PC-1024

Lab Name: Govt. PHC, Pavithram  
 Constituent Group: Chemistry III  
 PT Item: Lyophilized human serum based  
 Lab No: 12942  
 Date of Result Entered: 20/03/2023  
 Date of Report Published: 04/04/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants			Your Value	SDI	U
						CV	SD	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1878	91.48	7.46	6.82	102.35 mg/dL	1.59	0.31	
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1851	1.33	9.08	0.12	1.39 mg/dL	0.50	0.01	
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1761	110.93	5.23	5.80	118.64 mg/dL	1.33	0.28	
SDI Range											
				Interpretation							
Within -1.00 to +1.00				Excellent.							
Within ±1.01 to ±2.00				Good.							
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.							
Beyond ±3.0				Unacceptable performance. Action Signal.							

LAB ADDRESS :  
 Govt. PHC, Pavithram  
 PAVITHIRAM VILLAGE & POST  
 Pavithram  
 TAMILNADU606806

Coordinator Contact Details:  
 Email:clinqc@cmcvellore.ac.in  
 Contact Number: 0416-2283102



*Parvula Christudoss*  
 Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential  
 CMC EQAS does not sub contract any components  
 \*\*\*\*\* End of Report \*\*\*\*\*