



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - AUGUST 2023



PC-1024

Lab Name	Govt. Block PHC, Cheyur	Lab No	7613
Constituent Group	Chemistry III	Date of Result Entered :	18/08/2023
PT Item	Lyophilized human serum based	Date of Report Published :	01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonk (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	182 mg/dL	0.71	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonk (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.77 mg/dL	0.08	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonk (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	112 mg/dL	-0.39	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. Block PHC, Cheyur
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, TIRUPUR DISTRICT.
Cheyur
TAMILNADU641655

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****