

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. Block PHC, Piranmalai

Lab No 7176

Month March

Year 2023

Constituent Group Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

[All Analyser Result](#) [Print](#)

Date of Result Entered : 20/03/2023

Date of Report Published : 04/04/2023

Click on the analyte to view Graphical Data

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	SD	Your Value	SDI	L
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1878	91.48	7.46	6.82	129 mg/dL	5.5	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1851	1.33	9.08	0.12	1.4 mg/dL	0.58	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1761	110.93	5.23	5.80	142 mg/dL	5.36	0.28
<b>Interpretation</b>										
<b>SDI Range</b>										
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

Links

[Modify Views](#)

[Modify Methods](#)

[Submit Lab Results](#)

[View Lab Monthly Summary](#)

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. Block PHC, Pirannalai

Lab No 7176

Month April

Year 2023

Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 19/04/2023

Date of Report Published : 06/05/2023

Click on the analyte to view Graphical Data

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	CV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1832	119.45	6.04	7.21	133 mg/dL	1.88	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1820	2.16	8.09	0.18	1.6 mg/dL	-3.11	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1750	111.61	5.19	5.79	119 mg/dL	1.28	0.28
<b>Interpretation</b>										
<b>SDI Range</b>										
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

Links

[Modely Address](#)
[Faculty Address](#)
[Subord Lab Results](#)

[View Lab Monthly Summary](#)

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. Block PHC, Piranmalai  
 Lab No 7176  
 Month May  
 Year 2023  
 Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 20/05/2023

Date of Report Published : 05/06/2023

Sr.No	Analyte	Method : Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	183 mg/dL	0.23	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1854	1.76	7.48	0.13	1.8 mg/dL	0.30	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1798	114.60	5.08	5.83	112 mg/dL	-0.45	0.27
<b>SDI Range</b>				<b>Interpretation</b>						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

Links

[Modify Address](#)

[Modify Methods](#)

[Submit Lab Results](#)

[View Lab Monthly Summary](#)