



CALIBRATION REPORT

Report No : HBMS/CAL/23-24		Calibration Date : 15.09.2023
		Calibration valid : 14.03.2024
1.1.CUSTOMER DETAILS:		
Name and Address of the organization		M/s, GOVERNMENT PRIMARY HEALTH CENTRE , THURINJAPURAM, THIRUVANNAMALAI.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):		
EQUIPMENT NAME		CENTRIFUGE
i	Manufactured by	LOBOTECH
ii	Model	BPI-152
iii	Serial No	2011761

CALIBRATION WORKSHEET

LOCATION: LABORATORY

Tabulation For – Centrifuge

S.no	Unit under test IN Minutes	Observed speed tachometer in RPM	Deviation measured \pm RPM	Combined uncertainty in RPM in % rdg
1	1000	1000	0	1.12
2	1200	1201	1	1.23
3	1500	1502	2	1.56
4	2000	2001	1	1.175

Acceptable deviation is \pm 5 RPM it depends on Max Speed Range



MD Reg No: TN/CBE/MD42/00016



GSTIN: 33AAJFH1833M1ZB



99400 28355, 9751206106

88833 15662

No.42/1, Navaindia Road, Peelamedu,

CBE – 04, hbmedicalsystem@gmail.com