

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. Block PHC, Erisinampatti  
 Lab No: 7610  
 Month: August  
 Year: 2023  
 Constituent: Chemistry III  
 Group: Chemistry III

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 19/08/2023

Date of Report Published : 01/09/2023

S.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participant CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	175 mg/dL	-0.06	0.42
2	CREATININE III	James Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.67 mg/dL	-0.76	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	115.27 mg/dL	0.32	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

- Modify Address
- Modify Methods
- Submit Lab Results
- View Monthly Summary
- Home
- Change Password

*P. Ravi*

**MEDICAL OFFICER,**  
**GOVT. PRIMARY HEALTH CENTRE,**  
 Erisinampatti

