

DR. N.TAMILSELVI,MBBS.,  
BLOCK MEDICAL OFFICER,  
ELATHUR, TENKASI DISTRICT.



O/o PRIMARY HEALTH CENTRE ,  
ELATHUR 627803,  
TENKASI DISTRICT.

29.09.2023

## LEGAL IDENTITY CERTIFICATE

This is to certify that the following  
government primary health centre is functioning under the department of  
Public Health and Preventive Medicine, Government of Tamilnadu.

Name of the District	Name of the Health Unit District	Name of the Block	Name of the Primary Health Centre
TENKASI	TENKASI	ELATHUR	ELATHUR

This Certificate is issued for submitting the application  
for the "NABL Medical Entry Level Testing M(EL)T Labs Program"

Medical Officer  
*[Signature]*  
**MEDICAL OFFICER**  
Govt. Primary Health Centre  
ELATHUR - 627803  
Tirunelveli Dist.,  
D.O. Code No : MF-337



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - AUGUST 2023



PC-1024

Lab Name	Govt. Block PHC, Elathur	Lab No	7597
Constituent Group	Chemistry III	Date of Result Entered :	19/08/2023
PT Item	Lyophilized human serum based	Date of Report Published :	01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	166 mg/dL	-1.06	0.42
2	UREA III	Urease UV / GLDH III	Robonik ( Semi automatic / Fully Automatic )	174	28.29	15.20	4.30	26 mg/dL	-0.53	0.65
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	1.8 mg/dL	0.34	0.01
4	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	109 mg/dL	-1.04	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

## LAB ADDRESS :

Govt. Block PHC, Elathur

THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, TIRUNELVELI DISTRICT.

Elathur

TAMILNADU627803

Coordinator Contact Details:  
Email:clinqc@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

\*\*\*\*\* End of Report \*\*\*\*\*

*Sy 123*  
MEDICAL OFFICER  
Govt. Primary Health Centre  
ELATHUR - 627803  
Tirunelveli Dist.,  
D.O. Code No : MF-337