



CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23/10348		MEDICAL DEVICES		Page No:1 of 1
Issue Date	18-08-2023			
SRF No & Date	SRF/23/00318-0001 & 17-08-2023			
Receipt Date	17-08-2023			
Calibration Date	17-08-2023			
Calibration Due	16-08-2024			
Customer Name & Address				
GOVERNMENT PRIMARY HEALTH CENTRE, SINGARAPETTAI- 635307, KRISHNAGIRI DISTRICT.				
Details of Device Under Calibration (DUC)				
Description	: SEMI AUTO ANALYZER	Make & Model	: ROBONIK & PRIETEST TOUCH	
Range	: MULTI	Sr. No	: ATCD220818RBK	
Resolution	: MULTI	Identification No	: NA	
DUC Condition	: SATISFACTORY	Location	: LABORATORY	
Environmental Conditions & Calibration Procedure Details				
Environmental Details	Temperature:25.6 ° C	Relative Humidity	52% RH	
Calibration Procedure No	SBS/CP/MD/20	Calibration done at	ONSITE	
Reference Standards Details				
S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

ELECTRICAL SAFETY

RESULTS

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resistance >20MΩ	Measured values in MΩ 100	Uncertainty in % (±) 13.92
2	Earth Leakage <5000μAfor B,BF,CF	Measured values in μA 181	Uncertainty in % (±) 7.5
3	Enclosure Leakage <500μAfor B,BF,CF	Measured values in μA 209	Uncertainty in % (±) 7.3

REMARKS

- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

Authorised Signatory

(Calibration Engineer)
K.SATHYAMOORTHY

Technical Manager
C.SHANMUGARAJ

Chief Executive