

Govt. PHC, Purnavasi

Lab Name

12530

Lab No

June

Month

2023

Year

Chemistry III

Constituent Group

Click on the analyte to view Graphical Data

All Analyser Result

Date of Result Entered : 19/06/2023

Date of Report Published : 04/07/2023

- [Details About Robust Analysis](#)
- [Detail About Monthly Summary](#)
- [Detail about SDI](#)

SLNo	Analyte	Method / Principle Name	Analyser Name	No. of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1823	286.63	5.17	14.82	272 mg/dL	-0.99	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1848	2.57	6.64	0.17	2.4 mg/dL	-0.99	0.01
3	CHOLESTEROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1804	113.73	4.46	5.07	111 mg/dL	-0.54	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.

Type here to search