

Lab Name Govt. PHC, Panzavasal
Lab No 12830
Month May
Year 2023
Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 18/05/2023

Date of Report Published : 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonk (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	185 mg/dL	0.44	0.44
2	CREATININE III	Jaffes Kinebc - Alkaline Picrate III	Robonk (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.7 mg/dL	-0.45	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonk (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	112 mg/dL	-0.45	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within =1.01 to =2.00	Good.
Within =2.01 to =2.99	Accept with caution. Warning Signal.
Beyond =3.0	Unacceptable performance. Action Signal.