

**Lab Name** Govt. PHC, Punalvasal  
**Lab No** 12830  
**Month** March  
**Year** 2023  
**Constituent Group** Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 18/03/2023

Date of Report Published : 04/04/2023

SLNo	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1878	91.48	7.45	6.82	99 mg/dL	1.10	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate II	Robonik ( Semi automatic / Fully Automatic )	1851	1.33	9.08	0.12	1.1 mg/dL	-1.90	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1761	110.93	5.23	5.80	114 mg/dL	0.53	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.