

VIEW LAB MONTHLY SUMMARY

Lab Name: Govt PNC, Pappankulam Details About Robust Analysis
 Lab No: 12931 Detail About Monthly Summary
 Month: August Detail about SDI
 Year: 2023
 Constituent Group: Chemistry II

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered: 19/08/2023

Date of Report Published: 01/09/2023

| Sl. No. | Analyte | Method / Procedure Name | Analyser Name | No of Participants | CV | Precision | | True Value | SDI | SI |
|---------|---------------|--------------------------------------|---|--------------------|--------|-----------|------|-------------|-------|------|
| | | | | | | CV | SI | | | |
| 1 | GLUCOSE II | GOD-POD II | Rebortek (Semi automatic / Fully Automatic) | 1841 | 175.58 | 5.13 | 0.00 | 166.5 mg/dL | -1.01 | 0.42 |
| 2 | CREATININE II | Jaffee Kinetic - Alkaline Picrate II | Rebortek (Semi automatic / Fully Automatic) | 1852 | 1.76 | 6.77 | 0.12 | 1.96 mg/dL | 1.68 | 0.04 |
| 3 | CHOLESTROL II | CHOD-PAP II | Rebortek (Semi automatic / Fully Automatic) | 1764 | 193.80 | 4.07 | 4.64 | 119.6 mg/dL | 1.04 | 0.22 |

| SDI Range | Interpretation |
|----------------------|---|
| SDI < -1.00 to +1.00 | Excellent |
| SDI < ±1.01 to ±2.00 | Good |
| SDI < ±2.01 to ±2.99 | Accept with caution. Warning Signal |
| SDI < ±3.0 | Unacceptable performance. Action Signal |

| | | | |
|---------|-----------------|--------------------|--------------------------|
| Address | Modify Methods | Submit Lab Results | View Lab Monthly Summary |
| Summary | Change Password | Home | Logout |

MEDICAL OFFICER
G GOVT. PRIMARY HEALTH CENTRE,
PAPPANKULAM,
 Madathuruthy, P.O., Thrissur District