

LAB MONTHLY SUMMARY

Lab Name: Govt. Block PHC, D.Thurinjipatti
 Lab No: 7353
 Month: August
 Year: 2023
 Constituent: Chemistry III

Details About Robust Analysis
 Detail About Monthly Summary
 Detail about SDI

Click on the analyte to view Graphical Data

Date of Result Entered : 21/08/2023

Date of Report Published : 01/09/2023

Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
					CV	SD			
GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	172 mg/dL	-0.40	0.42
CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.9 mg/dL	1.18	0.01
CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	110 mg/dL	-0.82	0.22

SDI Range	Interpretation
SDI in -1.00 to +1.00	Excellent.
SDI in ±1.01 to ±2.00	Good.
SDI in ±2.01 to ±2.99	Accept with caution. Warning Signal.
SDI in ±3.0 and above	Unacceptable performance. Action Signal.

Home	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Summary	Change Password	Home	Sign-out

[Handwritten Signature]
 MEDICAL OFFICER
 Govt. Primary Health Centre
 D. THURUJIPATTY - 636 702
 Dharmapuri Dt.

MONTHLY SUMMARY

name	Govt. Block PHC, D.Thurinjipatti	Details About Robust Analysis
o	7353	Detail About Monthly Summary
	July	Detail about SDI
	2023	
stituent	Chemistry III	

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 20/07/2023

Date of Report Published : 07/08/2023

Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
					CV	SD			
LUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	101 mg/dL	-1.38	0.28
REATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	5.4 mg/dL	-1.23	0.02
HOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	114 mg/dL	0.21	0.22

SDI Range	Interpretation
-1.00 to +1.00	Excellent.
±1.01 to ±2.00	Good.
±2.01 to ±2.99	Accept with caution. Warning Signal.
d ±3.0	Unacceptable performance. Action Signal.

Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Summary	Change Password	Home	Sign-out

[Handwritten Signature]
MEDICAL OFFICER
 Govt. Primary Health Centre
 D. THURIJIPATTY - 635 702
 Dharmapuri Dt

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JUNE 2023



Govt. Block PHC, D.Thurinjipatti Lab No 7353
 Chemistry III Date of Result Entered : 20/06/2023
 Lyophilized human serum based Date of Report Published : 04/07/2023

Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	Participants		Your Value	SDI	U
					CV	SD			
COSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	296 mg/dL	0.63	0.69
ATNINE III	Jaffee Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.6 mg/dL	0.18	0.01
CESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	114 mg/dL	0.05	0.24

SDI Range	Interpretation
to +1.00	Excellent.
to ±2.00	Good.
to ±2.99	Accept with caution. Warning Signal.
	Unacceptable performance. Action Signal.

SS :
 PHC, D.Thurinjipatti
 I. OFFICER, D.THURINJIPATTI PHC, DHARMAPURI, DHARMAPURI DIST.
 II
 535202

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****

[Handwritten Signature]
MEDICAL OFFICER
 Govt. Primary Health Centre
 D. THURIJIPATTY - 635 702
 Dharmapuri - DL

ANALYSIS SUMMARY

Govt. Health Centre, B. Thuriappatti

Details About Robust Analysis

5333

Detail About Monthly Summary

Max

Detail about SDI

2023

Chemistry #

Click on the analyte to view Graphical Data

[All Analyser Results](#) [Print](#)

Date of Result Entered : 20/05/2023

Date of Report Published : 05/06/2023

AN	Method / Principle Name	Analyser Name	No of Participants	CV	Participants		Year Value	SDI	U
					CV	SD			
HE #	GLUC-POD #	Robosnik (Semi Automatic / Fully Automatic)	1834	180.88	5.19	5.39	170.5 mg/dl	-0.47	0.44
ONE #	Jaifex Krietic / Akabite Proxide #	Robosnik (Semi Automatic / Fully Automatic)	1884	1.78	7.48	0.13	1.91 mg/dl	1.14	0.01
TRGL #	CHOD-PAP #	Robosnik (Semi Automatic / Fully Automatic)	1798	114.89	5.08	5.83	113 mg/dl	-0.27	0.27

SI Range	Interpretation
to +1.00	Excellent.
to +2.00	Good.
to +3.00	Accept with caution. Warning Signal.
	Unacceptable performance. Action Signal.

Modify Methods	Submit Lab Results	View Lab Monthly Summary
Change Password	Home	Sign-out

D. Thuriappatti
 MEDICAL OFFICER
 Govt. Primary Health Centre
 B. THURIAPPATTI - 633 701
 Dharmapuri Dt.

MONTHLY SUMMARY

Name: Govt. Block PHC, D.Thurinjipatti
 Details About Robust Analysis
 ID: 7353
 Details About Monthly Summary
 Month: March
 Details about SDI
 Year: 2023
 Analyte: Chemistry III

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 20/03/2023

Date of Report Published : 04/04/2023

Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
					CV	SD			
GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.46	6.82	95 mg/dL	0.52	0.31
CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08	0.12	1.28 mg/dL	-0.41	0.01
CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1761	110.93	5.23	5.80	110 mg/dL	-0.16	0.28

SDI Range	Interpretation
n -1.00 to +1.00	Excellent.
n ±1.01 to ±2.00	Good.
n ±2.01 to ±2.99	Accept with caution. Warning Signal.
n ±3.0	Unacceptable performance. Action Signal.

Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Summary	Change Password	Home	Sign-out

[Signature]
MEDICAL OFFICER
 Govt. Primary Health Centre
 D. THURIJIPATTY - 635 702
 Dharmapuri Dt

B MONTHLY SUMMARY

Name	Govt. Block PHC, D.Thurinjipatti	Details About Robust Analysis
ID	7353	Detail About Monthly Summary
Month	December	Detail about SDI
Year	2022	
Subent	Chemistry III	

Click on the analyte to view Graphical Data

Date of Result Entered : 20/12/2022

Date of Report Published : 05/01/2023

Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
					CV	SD			
COSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1844	239.20	6.25	14.96	211 mg/dL	-1.88	0.70
ATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1839	3.91	8.24	0.32	4.1 mg/dL	0.59	0.02
LESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1766	114.94	5.41	6.21	110 mg/dL	-0.80	0.30

SDI Range	Interpretation
00 to +1.00	Excellent.
.01 to ±2.00	Good.
.01 to ±2.99	Accept with caution. Warning Signal.
>3.0	Unacceptable performance. Action Signal.

Modify Methods	Submit Lab Results	View Lab Monthly Summary
Change Password	Home	Sign-out