



CALIBRATION / TESTING REPORT			
CUSTOMER DETAILS : GOVERNMENT PRIMARY HEALTH CENTRE, KEELANATHAM, TIRUNELVELI, TAMILNADU-627353		CERTIFICATE NO	TICS/2023/KLNM/011
		EQUIPMENT NAME	CENTRIFUGE
		MANUFACTURE	REMI
		MODEL	C-852
CALIBRATION DATE	09/05/2023	SERIAL NO	ZDDN-10673
CALIBRATION DUE DATE	08/05/2024	INSTALLED DEPARTMENT	LAB

- 1. General maintenance required  Yes  No
- 2. Any Physical damage observed  Yes  No
- 3. Spare Replacement (if required)  Yes  No

TEST EQUIPMENT	MANUFACTURE	MODEL	SERIAL NUMBER
DIGITAL TACHOMETER	KUSAM MECO	KM2234 BL	S1049875

TEST REPORT				
				Temperature 25°C Humidity RH 55%

PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
ROTATION PER MINUTE	RPM	Speed 2	2236	± 5	PASS
		Speed 4	3207	± 5	PASS

\*The above mentioned instrument has been tested using standard manufacturer recommended protocols, using equipment having traceability to National/International standards.

ACCEPTABLE FOR USE  CORRECTIVE MAINTENANCE REQUIRED  REMOVED FROM USE

Calibrated By :

Authorized Sign



\*Electrical safety test result attached