



CALIBRATION / TESTING REPORT			
CUSTOMER DETAILS : GOVERNMENT PRIMARY HEALTH CENTRE, KEELANATHAM, TIRUNELVELI, TAMILNADU-627353		CERTIFICATE NO	TICS/2023/KLNM/014
		EQUIPMENT NAME	PIPETTE
		MANUFACTURE MODEL	THERMO SCIENTIFIC FINNPIPETTE
CALIBRATION DATE	09/05/2023	SERIAL NO	RW09696
CALIBRATION DUE DATE	08/05/2024	INSTALLED DEPARTMENT	LAB

- 1. Have more dust in machine outside Yes No
- 2. Any Physical damaged in machine Yes No
- 3. Machine inside Cleaning (if required) Yes No

TEST EQUIPMENT	MANUFACTURE	CERTIFICATE NUMBER	VALIDITY
ELECTRONIC SEMI MICRO BALANCE	SHIMADZU	TVCSP122/07/884-01	27-JUL-2023

TEST REPORT

Temperature 25°C
Humidity RH 55%

PARAMETERS	UNITS	SET VALUE	MEASURED VALUES	TOLERANCE	STATUS
VOLUME	µL	10	10.1	± 1	PASS
		50	50.5	±1	PASS
		100	100.4	± 1	PASS

*The above mentioned instrument has been tested using standard manufacturer recommended protocols, using equipment having traceability to National/International standards.

ACCEPTABLE FOR USE CORRECTIVE MAINTENANCE REQUIRED REMOVED FROM USE

Calibrated By :

Authorized Sign With Seal

