

CALIBRATION CERTIFICATE

Name and address of the customer

The Medical Officer
Government Primary health centre
Melaisivapuri -pudukkottai(dist)

Certificate no : ULR Number: CC275219000083790F
Date of calibration : Rmtl/01/319107848/13
Suggested calibration due : 06/03/2023
Customer po .reference.& date: : 06/03/2024

1.DESCRPTION OF DUC

Name Of The Duc: SEMIAUTO ANALYSER
Make: ROBONIK
Model : PRETEST TOUCH
Sl.No : ATcd1220321RBK
Range/Size : 230v

Receiving Condition Of Duc : Good
Ref. Document :As Per Manufactures Specification
Environmental Condition :25.3 Deg & 54.3% Rh
Uncertainty Of Measurement :As Per Report
Location : LAB

THE MASTERS ARE TRACEABLE TO NATIONAL/INTERNATIONAL STANDARDS AS PER ISO/IEC 17025.

2.DETAILED OF MASTER USED

Standard Instrument Used	Make/ Model	Id No	Traceable To/Certificate No	Validity
Decade Megohm	Vaisashika	Rml-230	Vaiseshika,Ambala Cantt/Ved/Nabl/18-19(Et)/92	1-6-2023
Multi Product Calibrator	Fluke	Rml-262	Wavecal,Blr/Cs-18/01-01	1-6-2023
Multi Product Calibrator (Ac/Dc Current Coil)	Transmille	Rml-252-1	Wavecal,Blr/Cs-18/69-01	1-6-2023
Leakage Current Tester	Htc/CI-2054	Rml-254-1	Rmtl/01/319107848	1-6-2023

CALIBRATION FIELD: Electro Technical (Voltage Calibration -Ac)

Sl.no	Range	Master readings v	Calibrated values v	Error observed v	Uncertainty v
1	0 v to 400v	230	231	1	0.055

CURRENT CALIBRATIONAc Current @50Hz

Sl no	Range	Master readings mA	Calibrated values mA	Error observed mA	Uncertainty mA
1	40Ma-100A	2	2.002	-0.002	0.045

Resistance calibration

Sl.no	Range	Master readings ohms	Calibrated values ohms	Error observed ohms	Uncertainty Ohms
1	40	39	39.01	0.001	0.007

4.remarks

- 4.1 uncertainty estimated for a confidence level of 95% at the k factor=2
 - 4.2the reported results are valid only for the condition of the inst./gauge received at the time of and under the condition of the calibration
 - 4.3 DUC -device under calibration
- Calibration label /status : Calibration result pass & fail are indicated



CALIBRATION ENGINEER



CALIBRATION SEAL



AUTHORIZED SIGN

This Report Refers Only To The Sample Submitted And May Not Be Reproduces In Full Or In Part, Without Written Permission From Roots Metrology & Testing Laboratory, Coimbatore, cell:9585840192