

Certificate No : PHCV/007



CALIBRATION CERTIFICATE

| Customer Details | Details of DUC | |
|---|--------------------|------------|
| URBAN PRIMARY HEALTH CENTRE VADASERY VANIGAR STREET VADASERY, NAGERCOIL, TAMILNADU 629001 | Description | CENTRIFUGE |
| | Manufactured by | NIL |
| | Model | NIL |
| | Serial No. | 5535 |
| | Identification No. | NIL |
| | Location | LAB |

| Environmental Conditions | | Calibration Details | |
|--------------------------|------|-------------------------|------------|
| Operating Temperature | 25°C | Date of Receipt | 08/02/2023 |
| | | Date of Calibration | 08/02/2023 |
| Humidity | 63 % | Date of Issue | 10/02/2023 |
| | | Due Date of calibration | 07/02/2024 |
| | | Calibrated At | Site |

Details of Calibration Standards

| Name | Manufacturer | Serial No. | Valid up to | Traceability |
|----------------------------|-----------------|---------------|-------------|---|
| Electrical safety analyser | Fluke / ESA 615 | 2244039 | 25-SEP-2023 | Accuver callibration laboratory CO LLC, Dubai |
| Digital Tacho Meter | Amprobe | BMT TACHO 031 | 30/09/2023 | Equation |

Performance Analysis

| Parameter | Set Value (RPM) | Measured Value(RPM) | Tolerance | Error Observed |
|-----------|-----------------|---------------------|-----------|----------------|
| RPM | 1000 | 1005 | ±5 % | 5 |
| | 2000 | 2006 | ±5 % | 6 |