

CALIBRATION CERTIFICATE

Name and address of the customer
The Medical Officer
Government Primary health centre
Melaisivapuri –pudukkottai(dist)

Certificate no : ULR Number: CC275219000083790F
Date of calibration : Rmtl/01/319107848/32
Suggested calibration due : 06/03/2023
Customer po .reference.& date: : 06/03/2024

1.DESRIPTION OF DUC

Name Of The Duc: pipette
Make: microlux
Model :
Sl.No :
Range/Size : 10 micro litter

Receiving Condition Of Duc : Good
Ref. Document : As Per Manufactures Specification
Environmental Condition : 25.3 Deg & 54.3% Rh
Uncertainty Of Measurement : As Per Report
Location : lab

THE MASTERS ARE TRACEABLE TO NATIONAL/INTERNATIONAL STANDARDS AS PER ISO/IEC 17025.

2.DETAILS OF MASTER USED

Standard Instrument Used	Make/ Model	Id No	Traceable To/Certificate No	Validity
Multi Product Calibrator	Fluke	Rml-262	Wavecal,Blr/Cs-18/01-01	1-6-2023
Micro balance	mettler toledo/XP26PC	QP-01	B115129039/CC-2523	1-6-2023
Universal Measuring Instrument With Sensor For Environmental Parameter	ALMEMO/2590-4AS	QP-10	H19070544/CC-2052	1-6-2023

CALIBRATION FIELD: WEIGHT Calibration

Sl.no	Range	Master readings micro litter	Calibrated values micro litter	Error observed micro litter	Uncertainty
1	10 micro litter	10	9.9	0.1	0.055

4.remarks

- 4.1 uncertainty estimated for a confidence level of 95% at the k factor=2
- 4.2the reported results are valid only for the condition of the inst./gauge received at the time of and under the condition of the calibration
- 4.3 DUC –device under calibration
- Calibration label /status : Calibration result pass & fail are indicated



CALIBRATION ENGINEER



CALIBRATION SEAL



AUTHORIZED SIGN

This Report Refers Only To The Sample Submitted
And May Not Be Reproduces In Full Or In Part,
Without Written Permission From Roots Metrology
& Testing Laboratory, Coimbatore, cell:9585840192