

Certificate No : PHCV/008



CALIBRATION CERTIFICATE

Customer Details	Details of DUC	
URBAN PRIMARY HEALTH CENTRE VADASERY VANIGAR STREET VADASERY, NAGERCOIL, TAMILNADU 629001	Description	SEMI AUTO ANALYZER
	Manufactured by	ROBONIK
	Model	PRIEST TOUCH
	Serial No.	NIL
	Identification No.	NIL
	Location	LAB

Environmental Conditions		Calibration Details	
Operating Temperature	25°C	Date of Receipt	08/02/2023
Humidity	63 %	Date of Calibration	08/02/2023
		Date of Issue	10/02/2023
		Due Date of calibration	07/02/2024
		Calibrated At	Site

Details of Calibration Standards

Name	Manufacturer	Serial No.	Valid up to	Traceability
Electrical safety analyser	Fluke / ESA 615	2244039	25-SEP-2023	Accuver calibration laboratory CO LLC, Dubai

Electrical Safety Test

Parameter	Measured Value	Tolerance	Remarks
Ground wire Resistance	0.150 Ω	<0.3 Ω	pass
insulation Resistance	OL	> 2 M Ω	pass
Chassis Leakage Current *NC	1.4 μ A	<100 μ A	Pass
Chassis Leakage Current *SFC	2.9 μ A	<500 μ A	Pass
Patient Leakage Current - B	1.2 μ A	<100 μ A	pass